

## ANNEX

EU TYPE EXAMINATION CERTIFICATE Nr. 049/2021/1217

## 1. Applicant

Coldes Textilien GmbH  
Goliplatz Am Alten Fließ 66  
50129 Bergheim  
Germany

## 2. Description

EN 13034:2005+A1:2009



Type PB[6]

EN 14126:2003+AC:2004



Type PB[6]-B

## 3. Materials and accessories

**Outer fabric**

- G-40gsm

**Cuff**

- Rib ref RC-001, color white

**Seam**

- Laminated fabric 40 gr. + seam Coats 120 (Coats)

## 4. Technical documentation

### Summary test results

EN 13034:2005+A1:2009

Outer fabric

G-40gsm

Method	Description	Result	Class
ISO 6330	Pre-treatment	/	/
ISO 15797	Pre-treatment	/	/
ISO 5077	Dimensional change of woven textile mat.	/	/
ISO 9073 4	Tear resistance trapezoidal	PASS	Class 1
EN ISO 13934 1	Tensile strength and elongation	PASS	Class 2
EN ISO 12947-2: 2016 (acc EN ISO 14325:2018)	Abrasion resistance	PASS	Class 2
EN 863	Puncture resistance	PASS	Class 1
EN 530 Methode 2 (acc EN ISO 14325:2004)	Abrasion resistance	/	/
EN 368 [EN ISO 6530] repellency o-xylene	Liquid repellency	PASS	Class 3
EN 368 [EN ISO 6530] repellency NaOH	Liquid repellency	PASS	Class 3
EN 368 [EN ISO 6530] repellency H2SO4	Liquid repellency	PASS	Class 3
EN 368 [EN ISO 6530] repellency butan-1-ol	Liquid repellency	PASS	Class 3
EN 368 [EN ISO 6530] penetration o-xylene	Liquid penetration	PASS	Class 3
EN 368 [EN ISO 6530] penetration NaOH	Liquid penetration	PASS	Class 3
EN 368 [EN ISO 6530] penetration H2SO4	Liquid penetration	PASS	Class 3
EN 368 [EN ISO 6530] penetration butan-1-ol	Liquid penetration butan-1-ol	PASS	Class 3

EN 14126:2003+AC:2004

Outer fabric

G-40gsm

Method	Description	Result	Class
ISO/DIS 22611	resistance to penetration by contaminated liquid aerosols	PASS	Class 3
EN ISO 7854 methode B - 30°C	Flex cracking resistance at -30°C	/	/
EN ISO13934 1	Tensile strength and elongation	PASS	Class 2
EN 13274-4 methode 3	Résistance à la flamme	/	/
ISO/FDIS 16604	penetration by infective agents bacteriophage test	PASS	Class 6
EN 530 Methode2	Abrasion resistance	/	/
ISO 9073 4	Tear resistance trapezoidal	PASS	Class 1
EN 368 penetration	Liquid penetration	PASS	Class 3
EN 368 [EN ISO 6530] repellency H2SO4	Liquid repellency	PASS	Class 3
ISO/FDIS 16603	penetration by infective agents synthetic blood test	PASS	Class 6
Annex A - EN14126	Test method for resistance to wet bacterial barrier penetration	PASS	Class 6
ISO 15797	Pre-treatment	/	/
EN 374-3 or EN ISO6529 A or B	Resistance to permeation by chemicals	/	/
EN ISO 7854 methode 2	Flex cracking resistance	/	/



ISO/DIS 22612	resistance to penetration by contaminated solid particles	PASS	Class 3
ISO 6330	Pre-treatment	/	/
EN 863	Puncture resistance	PASS	Class 1
EN ISO 12947-2: 2016 (acc EN ISO 14325:2018)	Abrasion resistance	PASS	Class 2

EN ISO 13688:2013 Outer fabric **G-40gsm**

Method	Description	Result	Class
ISO 3071 textile	pH	PASS	
ISO 3071 textile	pH	PASS	
EN 14362-1	Arylamines	PASS	
ISO 5077	Dimensional stability	/	/

EN ISO 13688:2013 Cuff **Rib ref RC-001, color white**

Method	Description	Result	Class
ISO 3071 textile	pH	PASS	
EN 14362-1	Arylamines	/	/

EN 13034:2005+A1:2009 Seam **Laminated fabric 40 gr. + seam Coats 120 (Coats)**

Method	Description	Result	Class
ISO 6330	Pre-treatment	/	/
ISO 15797	Pre-treatment	/	/
ISO 13935-2	Seam strength	PASS	Class 1

EN 14126:2003+AC:2004 Seam **Laminated fabric 40 gr. + seam Coats 120 (Coats)**

Method	Description	Result	Class
ISO 13935-2	Seam strength	PASS	Class 1
EN 374-3 or EN ISO6529 A or B	Resistance to permeation by chemicals	/	/
ISO 15797	Pre-treatment	/	/
ISO 6330	Pre-treatment	/	/

**Description/Picture of article**

Article **Coldes Infection Protection Gown**



*COLDES Textilien GmbH*

The above picture is a general picture of the article. Possible variations of the above article can be present in the technical file.

Note :

Any modification in material, design, or other technical features must be brought to the attention of the Notified Body.